

Comparison on Weighing Instrument Calibration between Eramed Kosovo and GDM Albania

Luljeta Disha¹, Defrim Bulku², Antoneta Deda¹

¹Sector of Mass and Density, General Directorate of Metrology, Albania

²Department of Physics, Faculty of Natural Science, Albania

Abstract: The General Directorate of Metrology (GDM) Albania, signed the agreement on collaboration with the Accredited Laboratory ERAMED N.T.SH from Kosovo, to conduct a proficiency testing (PT) in the field of calibration of weighing instrument. Usually testing and calibrating laboratories work with ISO/IEC 17025 standard, one of which requirements is that the laboratories should have quality control procedures for monitoring the validity of tests and calibrations undertaken. It is therefore important that such laboratories should regularly participate in a PT. A weighing instrument with capacity $Max = 520\text{ g}$ and resolution $d = 0.1\text{ mg}$ was the subject of this comparison. The measurements were performed on February 2017, at mass laboratory of GDM. This paper describes and analyzes the comparative results and uncertainties associated obtained by this two laboratories. The aim of this comparison is to verify the competence of ERAMED laboratory in the field of calibration of NAWI (non automatic weighing instrument.) The results are analyzed using normalized errors values (En).

Keywords: calibration, normalized error, proficiency testing, uncertainty, weighing instrument

I. Introduction

The mass laboratory of GDM is the national standard laboratory for mass in Albania. There are realized different Key Comparisons in which mass laboratory of GDM was involved, in the framework of IPA 2008 project supported by European Commission and some other supported by European Association of National Metrology Institutes (EURAMET). The degrees of equivalence resulting from these comparisons are within their respective standard uncertainties showing that the Key Comparison was a success for mass laboratory in Albania. Participation in a proficiency testing can validate the participating laboratory's measurement method, technical training, traceability of standards, and uncertainty budgets. The subject of this bilateral comparison was an electronic weighing instrument with $Max = 520\text{ g}$ and resolution $d = 0.1\text{ mg}$. This instrument has not been transported. Both laboratories agreed to perform the measurements at mass laboratory in Albania. This paper reports the results obtained by two laboratories in a proficiency testing scheme. The aim of the comparison was not only to compare measurement results of the participant with those of Pilot laboratory, but also to analyze measurement uncertainty, the choice of calibration points and to check the validity of quoted calibration measurement capabilities (CMC). The result is considered successful, if the value of the normalized error is $-1 \leq En \leq 1$. In this case the participating laboratory agrees with the reference value within the stated uncertainty. This paper is of interest for accredited or non-accredited laboratories for the calibration of NAWI as well for accreditation bodies.

MEASUREMENT INSTRUCTION

Initially, ERAMED mass laboratory of Kosovo announced its participation in the comparison of weighing instrument calibration and the mass laboratory of GDM accepted the role as the Pilot laboratory. Calibration will be performed over the full weighing range from zero to the maximum capacity. The test loads used for determining the errors of indication were: 0 g, 100g, 200 g, 300 g, 400 g and 500 g. Metrological and technical characteristics of the weighing instrument were provided before starting the proficiency testing [1]. Participants will use their standards procedures of calibration for which they will be validated and claim their CMC. The errors of indication and uncertainty associated to these errors were determined according to [2]. For acclimatization, it was proposed that ERAMED mass laboratory brings in advance the weights at mass laboratory in Albania. Standards weights used for calibration by Pilot laboratory have uncertainty according to CMC. The comparison measurements were carried out on February 2017 and the time schedule is reported in Table 1.

Table 1 Time schedule of comparison

Time - start	Time - end	Time periods Working days	Action
01 February 2017	14 February 2017	14 days	Preparation of technical protocol PL GDM
14 February 2017	15 February 2017	2days	Measurements from ERAMED Laboratory
16 February 2017	22 February 2017	7 days	Measurements from Pilot Laboratory
24 February 2017	30 April 2017	35 days	Analyses of results and final report from Pilot Laboratory

In order to be in the same conditions of calibration performance, both laboratories will do the adjustment of the weighing instrument before calibration according to the user manual.

For the calibration of the balance the Pilot laboratory used direct comparison method with standard weights having nominal mass and accuracy class presented in the Table 2. The standard weights used by ERAMED laboratory are shown in Table 3

Table 2 Standard weights used for calibration by GDM mass laboratory

Nominal value	Manufacturer	Serial Number	Identification	Class	Uncertainty (U)	Traceability /Certificate
1 g – 500 g	HAIGIS	7259	LM-07/03	E2	(4.0 - 95) µg	LM-1116G-223 GDM

Table 3 Standard weights used for calibration by ERAMED mass laboratory

Nominal value	Manufacturer	Serial Number	Identification	Class	Uncertainty (U)	Traceability/ Certificate
1 mg – 1 kg	SARTORIUS	26629505	Inv. Nr-026	E2	(3.0 - 500) µg	BoM Macedonia 12.2016

Standards weights used for calibration by Pilot Laboratory have the form and accuracy according to [3] and are traceable to the primary mass laboratory of GDM. Repeatability was determined at 500g and 200g. For each load, ten measurements were performed. The test of eccentricity is performed at 200g. The error of indication and uncertainty associated to these errors were determined according to equation (1) and (2) respectively. For loading the balance, both laboratories have opted to load the balance only once, increasing, by steps with unloading between the separate steps, corresponding to the majority of uses of the instruments for weighing single loads.

$$E_j = I_j - m_{ref} \tag{1}$$

I_j - the indication of the balance
 m_{ref} - is conventional mass of the reference weights

$$U(E) = 2u(E) = \sqrt{\frac{d_0^2}{12} + \frac{d_L^2}{12} + s^2(I) + u_{ecc}^2 + u^2(m_{ref})} \tag{2}$$

The signification of the terms from the formula (2) is explained in Fig.1 [4].




Figure1 Ishikawa diagram of uncertainty components in the calibration of the balance

II. Analyses And Results Of Comparison

After completion of the measurements, the measurements results of participant laboratory were sent to the scheme coordinator in the required format of the proficiency test protocol, in a period of 7 working days. Then, the Pilot laboratory performed the analysis of results. Consistency of the test is considered satisfactory when the probably test of the results for participant χ^2 is fulfilled [5].

According to the analyses of Pilot laboratory the condition $\Pr\{\chi^2(v) > \chi^2_{obs}\} < 0.05$ is fulfilled for each load. Depending on the evaluation procedure used, the reference value is defined as the weighted mean of Pilot laboratory measurements according to equation (3) [5].

$$\Delta m_{ref} = \frac{\sum_{i=1}^N \frac{\Delta m_{c-p}}{u(\Delta m_{c-p})^2}}{\sum_{i=1}^N \frac{1}{u(\Delta m_{c-p})^2}} \quad (3)$$

where $\overline{\Delta m_c} = \Delta m_c = \text{Ref}$

In order to carry out consistency checks of the results the uncertainty in the reference value has been calculated also according to the methods described by [5] with equation (4).

$$u(\Delta m_{c-ref}) = \frac{1}{\sum_{i=1}^N \frac{1}{u(\Delta m_{c-p})^2}} \quad (4)$$

where: $U(\overline{\Delta m_w}) = U(\Delta m_{ref}) = U_{Ref}$

In Table 4 below are presented all the components of uncertainty reported by the participant laboratory and by the Pilot laboratory.

Table 4 The components of uncertainty reported by ERAMED and Pilot laboratory

(Standard) uncertainty contributions (mg)			
Uncertainty component	Symbol	GDM	ERAMED
Repeatability	(s)	x	x
Resolution (0 load)	$u(R_0)$	x	x
Resolution (L load)	$u(R_L)$	x	x
Uncertainty due to eccentricity effect	$u(\Delta m_{ecc})$	x	x
Uncertainty of the indication	$u(I)$	x	x
Uncertainty due to mass standards	$u(\Delta m_c)$	x	x
Uncertainty due to drift of mass standards	$u(\Delta m_D)$	x	x
Uncertainty due to buoyancy correction	$u(\Delta m_B)$	x	x
Uncertainty due to convection effect	$u(\Delta m_{conv})$	x	-
Standard uncertainty of the error of indication $u_c (k=1)$	$u(E)$	x	x

Table 5 presents the errors of indication "E", calculated according to formula (1) and the expanded uncertainty "U" (k = 2) associated with these errors, calculated according to formula (2), reported by Pilot Laboratory and ERAMED laboratory

Table 5 The errors of indication "E" and expanded uncertainty "U" reported by GDM and ERAMED laboratory

Load (g)	Laboratory			
	GDM		ERAMED	
	E (g)	U(g)	E (g)	U(g)
0	0.00000	0.00010	0.00000	0.00012
100	0.00050	0.00011	0.00060	0.00024
200	0.00109	0.00013	0.00120	0.00042
300	0.00159	0.00018	0.00180	0.00095
400	0.00223	0.00022	0.00250	0.00123
500	0.00310	0.00026	0.00360	0.00164

1.1 Normalized errors of the reported values and degrees of equivalence for participant laboratory

A tool often used in analyzing the results from interlaboratory comparisons is the normalized error E_n , which takes into account both the result and its uncertainty. The normalized error is calculated according to equation (5) [5]:

$$E_n = \left| \frac{(\Delta m_{ci} - \overline{\Delta m_{c-ref}})}{\sqrt{(u(\Delta m_{ci})^2 + u(\overline{\Delta m_{c-ref}})^2)}} \right| \quad (5)$$

The result is considered successful, if the value of the normalized error is $-1 \leq E_n \leq 1$. In Table 6 are presented the normalized errors E_n of the reported values and degrees of equivalence for participant laboratory. In this table the term $\Delta m_c (deq_A)$ is calculated as difference between participant values and pilot laboratory reference values and expresses the degree of equivalence for participant laboratory. In the other hand, $\Delta U (deq_A)$ is the term related to uncertainty and calculated also as a difference against reference uncertainty value.

Table 6 Normalized errors E_n and degrees of equivalence for participant laboratory

Nominal value	E_n ERAMED	$\Delta m_c (deq_A)$ (g)	$\Delta U (deq_A)$ (g)
500 g	0.301	0.00050	0.00162
400 g	0.216	0.00027	0.00121
300 g	0.217	0.00021	0.00093
200 g	0.250	0.00011	0.00040
100 g	0.379	0.00010	0.00021
0 g	0.000	0.00000	0.00007

From Table 6 it is visible that the values of normalized error are lower than 1 for participant laboratory and in this case it can be said that the measurements results of participant laboratory are consistent with the reference values. Figure 2 presents the reference values of Pilot laboratory and participant measurements results with corresponding uncertainties.




Figure 2 Reference values of Pilot laboratory and participant measurements results with corresponding uncertainties

III. Conclusions

A bilateral comparison in the field of calibration of weighing instruments between GDM, Albania and ERAMED, Kosovo was performed. Calibration laboratories announced their participation in a comparison of a NAWI, having Max 520g and $d = 0.1$ mg. In this case the participating laboratory agrees with the reference value within the stated uncertainty. If this laboratory has any unsuccessful results, i.e. $E_n < -1$ or $E_n > 1$, it is expected that the laboratory investigates the reason for the disagreement and implements corrective action. From this point of view, it can be concluded that bilateral comparison between the GDM mass laboratory and ERAMED mass laboratory was successful. Consequently the performance of the mass laboratory ERAMED was satisfactory, validating in this way its technical competence in the field of calibration of weighing instruments. In Albania, Proficiency Testing in the field of calibration of non automatic weighing instruments is a new subject.

References

- [1] OIML Recommendation R 76-1 "General requirements for metrological and technical characteristics" of NAWI. 2006
- [2] EURAMET, cg -18 "Guidelines on the Calibration of Non-automatic Weighing Instruments", Version 4.0 /2015
- [3] Weights of classes E1, E2, F1, F2, M1, M1-2, M2, M2-3 and M3, International recommendation OIML R111-1, OIML, (Paris), 2004
- [4] Adriana Valcu, "Between Parallel Mirrors – An Electronic Weighing Instrument", Proc. 9th International symposium on advanced topics in electrical engineering, Bucuresti, 2015.
- [5] Cox M. G. 2002, Metrologia **39** 589-595

Luljeta Disha. "Comparison on Weighing Instrument Calibration between Eramed Kosovo and GDM Albania." IOSR Journal of Applied Physics (IOSR-JAP) 9.4 (2017): 60-63.